



# CERTIFICATE OF ACCREDITATION

**ANSI-ASQ National Accreditation Board**

500 Montgomery Street, Suite 625, Alexandria, VA 22314, 877-344-3044

This is to certify that

**InterTech Development Company**  
**7401 North Linder Avenue**  
**Skokie, IL 60077**

has been assessed by ANAB  
and meets the requirements of international standard

**ISO/IEC 17025:2005**

while demonstrating technical competence in the field of

**CALIBRATION**

Refer to the accompanying Scope of Accreditation for information regarding the types of calibrations to which this accreditation applies.

L2197

Certificate Number

  
ANAB Approval

Certificate Valid: 12/07/2017-12/18/2020  
Version No. 001 Issued: 12/07/2017



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

### InterTech Development Company

7401 North Linder Avenue  
Skokie, IL 60077  
Steven Nowak (847) 679-3377 ext. 229

### CALIBRATION

Valid to: **December 18, 2020**

Certificate Number: **L2197**

#### Mass

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pneumatic Flow	(1 to 18) sccm	1.2% of reading	Laminar Flow System

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

#### Notes:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. L2197.



Vice President

